

**Notice of Allowability**

Application No.

10/577,108

Applicant(s)

NEDERPELT, CHRISTIANUS  
GERARDUS LAMBERT

Examiner

Hoon Song

Art Unit

2882

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☐ This communication is responsive to \_\_\_\_.
2. ☒ The allowed claim(s) is/are 1-10.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some\* c) ☐ None of the:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
3. ☒ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
- (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
- 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_.
- (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- |  |   |
|--|---|
| 1. <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                                | 5. <input type="checkbox"/> Notice of Informal Patent Application                     |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                       | 6. <input type="checkbox"/> Interview Summary (PTO-413),<br>Paper No./Mail Date ____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO/SB/08),<br>Paper No./Mail Date ____     | 7. <input type="checkbox"/> Examiner's Amendment/Comment                              |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit<br>of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance  |
|  | 9. <input type="checkbox"/> Other ____.   |

## **DETAILED ACTION**

### ***Allowable Subject Matter***

Claims 1-10 are allowed over prior art.

The following is an examiner's statement of reasons for allowance:

Regarding claims 1-5 , the prior art fails to teach an x-ray examination apparatus having image processing means to process at least a part of a first image with a grey-level compression function in order to compress the first image into a second image and control means to compute an average grey value of pixels of the second image and to provide a dose control signal to an x-ray source in order to control the x-ray beam, where the dose control signal being determined from a comparison between the computed average grey level value and a pre-stored reference value as claimed in independent claim 1.

Regarding claims 6-10, the prior art fails to teach a method of controlling an output of an x-ray source by processing at least a part of a first x-ray image with a grey-level compression function in order to compress the first x-ray image into a second x-ray image, comparing a computed average value of pixel of the second x-ray image and a pre-stored reference value and providing a dose control signal to the x-ray source, the dose control signal being determined from the comparison as claimed in independent claim 6

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably

Art Unit: 2882

accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Horbaschek (US 5,509,044) teaches a medical diagnostic system having an automatic exposure unit for optimizing signal acquisition for radiation exposure control but fails to teach the allowable subject matter above.


Fujiwara (US 5388138) teaches an x-ray controller for controlling a dose of x-ray but fails to teach the allowable subject matter above.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Hoon Song whose telephone number is (571) 272-2494. The examiner can normally be reached on 9:30 AM - 7 PM, Monday - Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward Glick can be reached on (571) 272 - 2490. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 2882

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Hoon Song  
Primary Examiner  
Art Unit 2882

5/3/2007